

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/210183328(Expiry date : 09 Mar 2023)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Jan 2023	:				
02 Jan 2023	:				
03 Jan 2023	:				
04 Jan 2023	8 : 30	4.5	3.1	5.7	S-H
05 Jan 2023	8 : 30	4.3	3.1	5.2	Y.N
06 Jan 2023	10 : 00	4.7	3.1	5.3	S-H
07 Jan 2023	9 : 00	3.5	3.1	5.7	S-H
08 Jan 2023	:				
09 Jan 2023	:				
10 Jan 2023	10 : 00	3.9	3.2	5.4	S-H
11 Jan 2023	10 : 30	4.6	3.5	5.4	S-H
12 Jan 2023	9 : 00	4.5	3.5	5.5	S-H
13 Jan 2023	11 : 30	4.9	3.5	5.4	R-Y
14 Jan 2023	9 : 00	3.7	3.6	5.2	S-H
15 Jan 2023	:				
16 Jan 2023	9 : 15	5.0	3.5	5.3	S-H
17 Jan 2023	14 : 00	3.6	3.4	3.9	S-H
18 Jan 2023	15 : 00	3.4	3.3	5.2	S-H
19 Jan 2023	9 : 00	4.5	3.2	5.2	S-H
20 Jan 2023	9 : 00	3.5	3.2	5.8	S-H
21 Jan 2023	9 : 30	4.5	3.2	5.2	Y.N
22 Jan 2023	:				
23 Jan 2023	9 : 30	5.5	3.2	6.3	S-H
24 Jan 2023	9 : 30	5.4	4.1	5.8	S-H
25 Jan 2023	14 : 30	3.9	3.1	5.1	S-H
26 Jan 2023	8 : 40	4.0	3.7	4.4	R-Y
27 Jan 2023	9 : 00	3.7	3.7	4.5	S-H
28 Jan 2023	9 : 00	5.1	3.7	5.7	S-H
29 Jan 2023	:				
30 Jan 2023	9 : 00	5.1	3.7	5.5	S-H
31 Jan 2023	10 : 00	4.9	3.7	5.6	S-H