

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/221808077(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Oct 2023	:				
02 Oct 2023	9 : 30	3.9	3.9	5.9	S-H
03 Oct 2023	10 : 30	3.9	3.9	5.8	S-H
04 Oct 2023	9 : 00	3.9	3.9	5.8	S-H
05 Oct 2023	15 : 30	3.9	3.9	5.8	Y.N
06 Oct 2023	10 : 30	3.9	3.9	5.8	S-H
07 Oct 2023	9 : 15	3.9	3.9	5.7	S-H
08 Oct 2023	:				
09 Oct 2023	:				
10 Oct 2023	10 : 00	3.9	3.8	5.8	S-H
11 Oct 2023	9 : 00	3.9	3.8	5.9	R-Y
12 Oct 2023	14 : 00	3.8	3.8	5.8	S-H
13 Oct 2023	8 : 45	3.8	3.8	5.7	S-H
14 Oct 2023	:				
15 Oct 2023	:				
16 Oct 2023	9 : 30	3.8	3.8	5.7	S-H
17 Oct 2023	10 : 00	3.9	3.8	5.7	S-H
18 Oct 2023	10 : 00	3.8	3.8	5.7	S-H
19 Oct 2023	9 : 00	3.8	3.8	5.7	S-H
20 Oct 2023	8 : 45	3.8	3.8	5.7	S-H
21 Oct 2023	8 : 30	3.8	3.8	5.6	S-H
22 Oct 2023	:				
23 Oct 2023	8 : 30	4.0	3.8	5.8	S-H
24 Oct 2023	8 : 30	4.0	4.0	5.7	S-H
25 Oct 2023	8 : 45	4.0	3.9	5.8	S-H
26 Oct 2023	8 : 50	4.0	4.0	5.8	S-H
27 Oct 2023	8 : 55	4.0	3.9	6.0	S-H
28 Oct 2023	9 : 00	4.0	3.9	5.7	S-H
29 Oct 2023	:				
30 Oct 2023	9 : 40	4.0	3.9	5.8	S-H
31 Oct 2023	14 : 00	4.0	3.9	5.9	S-H