

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/221808077(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Jan 2024	:				
02 Jan 2024	:				
03 Jan 2024	:				
04 Jan 2024	8 : 45	3.9	3.8	5.7	S-H
05 Jan 2024	14 : 00	3.8	3.7	5.5	S-H
06 Jan 2024	9 : 10	3.8	3.8	5.4	S-H
07 Jan 2024	:				
08 Jan 2024	:				
09 Jan 2024	8 : 45	3.8	3.8	5.9	S-H
10 Jan 2024	8 : 30	3.8	3.8	5.4	S-H
11 Jan 2024	9 : 45	3.8	3.8	5.5	S-H
12 Jan 2024	10 : 30	4.0	3.8	5.6	S-H
13 Jan 2024	12 : 00	4.2	3.8	5.6	S-H
14 Jan 2024	:				
15 Jan 2024	14 : 30	3.9	3.9	5.5	S-H
16 Jan 2024	10 : 30	3.9	3.9	5.4	S-H
17 Jan 2024	9 : 00	3.9	3.9	5.3	S-H
18 Jan 2024	8 : 50	3.9	3.8	5.4	S-H
19 Jan 2024	10 : 15	3.9	3.8	5.5	S-H
20 Jan 2024	9 : 00	3.9	3.8	5.6	S-H
21 Jan 2024	:				
22 Jan 2024	10 : 00	3.9	3.8	5.5	S-H
23 Jan 2024	11 : 00	3.9	3.8	5.7	S-H
24 Jan 2024	9 : 00	3.8	3.8	5.7	S-H
25 Jan 2024	9 : 00	3.8	3.8	5.7	S-H
26 Jan 2024	8 : 30	3.8	3.8	5.5	Y.N
27 Jan 2024	8 : 45	3.8	3.8	5.5	Y.N
28 Jan 2024	:				
29 Jan 2024	9 : 30	3.8	3.8	5.5	Y.N
30 Jan 2024	10 : 10	3.9	3.8	5.3	Y.N
31 Jan 2024	9 : 40	3.9	3.8	5.6	Y.N

Shunsuke Hitosugi 07/Feb/2024