

# Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/221808077(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Feb 2024	8 : 45	3.9	3.8	5.6	R.Y
02 Feb 2024	10 : 45	3.9	3.8	5.6	R.Y
03 Feb 2024	9 : 20	3.8	3.8	5.6	R.Y
04 Feb 2024	:				
05 Feb 2024	9 : 00	3.8	3.8	5.5	S.H
06 Feb 2024	10 : 00	3.8	3.8	5.6	S.H
07 Feb 2024	9 : 10	3.8	3.8	5.6	S.H
08 Feb 2024	9 : 15	3.8	3.8	5.6	S.H
09 Feb 2024	10 : 30	3.8	3.7	5.5	S.H
10 Feb 2024	8 : 45	3.8	3.7	5.5	S.H
11 Feb 2024	:				
12 Feb 2024	:				
13 Feb 2024	8 : 40	3.8	3.8	5.3	S.H
14 Feb 2024	8 : 30	3.8	3.8	5.1	S.H
15 Feb 2024	9 : 15	3.8	3.7	5.4	S.H
16 Feb 2024	9 : 00	3.8	3.8	5.4	S.H
17 Feb 2024	8 : 45	3.8	3.8	5.5	S.H
18 Feb 2024	:				
19 Feb 2024	11 : 15	3.8	3.7	5.7	S.H
20 Feb 2024	8 : 50	3.8	3.7	5.8	S.H
21 Feb 2024	9 : 00	3.8	3.7	5.9	S.H
22 Feb 2024	8 : 40	3.7	3.7	5.8	S.H
23 Feb 2024	:				
24 Feb 2024	9 : 00	3.8	3.7	5.6	S.H
25 Feb 2024	:				
26 Feb 2024	11 : 00	4.1	3.7	5.4	S.H
27 Feb 2024	9 : 00	3.7	3.1	5.2	S.H
28 Feb 2024	8 : 45	3.7	3.7	5.4	S.H
29 Feb 2024	9 : 00	3.7	3.7	5.3	S.H

Shunsuke Hitosugi 02 / Mar / 2024