

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/221808077(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Apr 2024	11 : 00	3.0	3.0	4.6	S.H
02 Apr 2024	9 : 30	3.0	3.0	4.5	S.H
03 Apr 2024	8 : 40	3.0	3.0	4.5	R.Y
04 Apr 2024	8 : 40	3.0	3.0	4.6	Y.N
05 Apr 2024	8 : 40	3.0	3.0	4.6	R.Y
06 Apr 2024	14 : 15	3.0	3.0	4.5	R.Y
07 Apr 2024	:				
08 Apr 2024	8 : 30	3.1	3.0	4.3	S.H
09 Apr 2024	10 : 30	3.1	3.0	4.8	S.H
10 Apr 2024	8 : 30	3.0 3.2	3.0	4.8	Y.N
11 Apr 2024	8 : 30	3.2	3.1	4.5	S.H
12 Apr 2024	14 : 00	3.1	3.0	4.7	S.H
13 Apr 2024	9 : 55	3.2	3.1	4.6	R.Y
14 Apr 2024	:				
15 Apr 2024	8 : 30	3.1	3.1	4.7	Y.N
16 Apr 2024	9 : 30	3.2	3.1	4.8	S.H
17 Apr 2024	9 : 00	3.0	3.0	4.6	S.H
18 Apr 2024	8 : 30	3.1	3.0	4.6	S.H
19 Apr 2024	9 : 00	3.0	3.0	4.6	R.Y
20 Apr 2024	11 : 00	3.6	3.0	4.4	Y.N
21 Apr 2024	:				
22 Apr 2024	9 : 00	3.4	3.0	5.1	S.H
23 Apr 2024	9 : 00	3.5	3.4	5.0	Y.N
24 Apr 2024	8 : 30	3.5	3.4	5.0	Y.N
25 Apr 2024	8 : 30	3.4	3.4	5.0	S.H
26 Apr 2024	8 : 30	3.4	3.4	5.1	S.H
27 Apr 2024	8 : 30	3.4	3.4	5.1	S.H
28 Apr 2024	:				
29 Apr 2024	:				
30 Apr 2024	12 : 00	4.1	3.4	5.2	S.H

* error
10-Apr-2024 Y.N

Investigational Product Manager's Signature: Shunsuke Hitosugi

Date: 07/ May/2024