

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/221808077(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Jun 2024	8 : 50	3.2	3.1	4.9	R-Y
02 Jun 2024	:				
03 Jun 2024	8 : 30	3.2	3.1	5.0	Y.N
04 Jun 2024	11 : 40	3.2	3.1	5.0	R-Y
05 Jun 2024	9 : 00	3.3	3.1	5.0	S-H
06 Jun 2024	9 : 30	3.5	3.1	5.0	S-H
07 Jun 2024	13 : 30	3.5	3.3	4.9	S-H
08 Jun 2024	8 : 30	3.3	3.2	4.9	S-H
09 Jun 2024	:				
10 Jun 2024	8 : 40	3.2	3.3	4.7	Y.N
11 Jun 2024	10 : 16	3.3	3.2	4.7	Y.N
12 Jun 2024	10 : 15	3.3	3.2	4.7	R-Y
13 Jun 2024	9 : 30	3.4	3.2	4.7	S-H
14 Jun 2024	9 : 00	3.4	3.2	4.9	S-H
15 Jun 2024	9 : 00	3.5	3.4	4.9	S-H
16 Jun 2024	:				
17 Jun 2024	11 : 10	4.0	2.9	5.1	R-Y
18 Jun 2024	10 : 15	3.1	3.0	4.6	R-Y
19 Jun 2024	9 : 10	3.2	3.0	4.7	R-Y
20 Jun 2024	9 : 30	3.1	3.0	4.7	S-H
21 Jun 2024	9 : 40	3.2	3.0	5.3	S-H
22 Jun 2024	9 : 00	3.0	2.9	4.5	S-H
23 Jun 2024	:				
24 Jun 2024	8 : 30	3.1	2.9	4.6	S-H
25 Jun 2024	8 : 30	3.9	3.0	5.7	S-H
26 Jun 2024	8 : 30	4.0	3.0	5.7	S-H
27 Jun 2024	8 : 40	3.9	3.8	5.6	Y.N
28 Jun 2024	9 : 20	3.9	3.9	5.6	S-H
29 Jun 2024	8 : 30	3.9	3.9	5.7	S-H
30 Jun 2024	:				

Investigational Product Manager's Signature: Shunsuke Hitosugi

Date: 03/Jul/2024