

# Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/221808077(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Aug 2024	8 : 30	3.9	3.7	5.8	S.H
02 Aug 2024	14 : 00	4.2	3.6	5.6	S.H
03 Aug 2024	8 : 30	3.9	3.8	5.6	S.H
04 Aug 2024	:				
05 Aug 2024	8 : 30	3.9	3.8	5.8	S.H
06 Aug 2024	8 : 30	4.0	3.8	5.7	Y.N
07 Aug 2024	8 : 30	4.1	3.9	5.8	S.H
08 Aug 2024	8 : 30	4.1	4.0	5.8	S.H
09 Aug 2024	8 : 30	4.0	3.8	5.7	S.H
10 Aug 2024	9 : 00	4.0	3.8	5.8	S.H
11 Aug 2024	:				
12 Aug 2024	:				
13 Aug 2024	8 : 30	3.8	3.6	5.8	S.H
14 Aug 2024	8 : 30	3.8	3.6	6.0	S.H
15 Aug 2024	8 : 30	3.8	3.7	6.3	Y.N
16 Aug 2024	9 : 00	3.7	3.6	6.3	S.H
17 Aug 2024	8 : 30	3.8	3.6	5.5	Y.N
18 Aug 2024	:				
19 Aug 2024	8 : 30	3.7	3.5	3.7	S.H
20 Aug 2024	8 : 30	3.7	3.4	6.2	S.H
21 Aug 2024	8 : 30	4.0	3.9	6.0	S.H
22 Aug 2024	9 : 00	3.9	3.9	5.7	R.Y
23 Aug 2024	9 : 00	3.9	3.9	5.7	Y.N
24 Aug 2024	8 : 30	3.9	3.8	5.7	Y.N
25 Aug 2024	:				
26 Aug 2024	8 : 30	4.0	3.8	5.9	S.H
27 Aug 2024	10 : 30	4.0	3.8	5.9	S.H
28 Aug 2024	12 : 30	4.0	3.8	5.7	Y.N
29 Aug 2024	8 : 30	4.0	3.8	5.7	S.H
30 Aug 2024	8 : 30	4.0	3.9	6.1	S.H
31 Aug 2024	9 : 00	4.1	3.9	5.7	S.H

Investigational Product Manager's Signature: Shunsuke Hitosugi

Date: 02 / Sep / 2024