

* 2024.10.15 S-H

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical departmentAcceptable Temperature Range : 2°C to 8°CRefrigerator/Freezer Name: ER-49CThermometer Device/ID: FS Traceable/221808077 (Expiry date : 05 Oct 2024) * 230735256 (Expiry date 01 Nov 2025)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Nov 2024	9 : 30	3.9	3.7	6.7	S-H
02 Nov 2024	8 : 45	3.9	3.8	6.7	S-H
03 Nov 2024	:				
04 Nov 2024	:				
05 Nov 2024	8 : 45	4.0	3.5	6.9	S-H
06 Nov 2024	9 : 00	4.0	3.7	6.7	S-H
07 Nov 2024	11 : 00	4.0	3.7	6.8	Y.N
08 Nov 2024	10 : 00	3.8	3.7	6.8	S-H
09 Nov 2024	12 : 00	3.8	3.1	6.8	S-H
10 Nov 2024	:				
11 Nov 2024	8 : 30	4.0	3.8	7.0	S-H
12 Nov 2024	9 : 30	3.8	3.8	7.2	S-H
13 Nov 2024	10 : 00	4.0	3.7	6.7	S-H
14 Nov 2024	8 : 30	3.9	3.7	6.8	S-H
15 Nov 2024	9 : 00	4.1	3.8	6.9	S-H
16 Nov 2024	9 : 20	3.9	3.8	7.0	Y.N
17 Nov 2024	:				
18 Nov 2024	8 : 30	4.0	3.8	7.1	S-H
19 Nov 2024	9 : 00	4.0	3.8	7.0	S-H
20 Nov 2024	11 : 00	4.1	3.8	6.7	S-H
21 Nov 2024	12 : 30	4.1	3.8	6.9	S-H
22 Nov 2024	10 : 20	4.2	3.9	6.9	Y.N
23 Nov 2024	:				
24 Nov 2024	:				
25 Nov 2024	9 : 00	4.1	3.8	6.9	S-H
26 Nov 2024	9 : 00	4.1	3.9	6.7	R-Y
27 Nov 2024	9 : 55	4.0	3.9	6.8	R-Y
28 Nov 2024	14 : 30	4.1	3.9	7.1	S-H
29 Nov 2024	10 : 00	4.1	3.9	6.7	S-H
30 Nov 2024	8 : 30	4.1	3.9	6.8	S-H

Investigational Product Manager's Signature: Shunsuke Hironaga;Date: 02-Dec-2024