

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/230735256(Expiry date : 01 Nov 2025)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Mar 2025	9 : 00	3.4	3.1	6.3	S.H
02 Mar 2025	:				
03 Mar 2025	8 : 30	3.2	3.0	6.5	S.H
04 Mar 2025	10 : 00	3.1	2.9	6.4	S.H
05 Mar 2025	9 : 00	3.7	2.6	6.5	S.H
06 Mar 2025	9 : 00	3.8	2.4	6.8	S.H
07 Mar 2025	10 : 00	3.5	3.3	6.3	S.H
08 Mar 2025	10 : 00	3.3	3.2	6.1	S.H
09 Mar 2025	:				
10 Mar 2025	10 : 00	3.3	3.2	6.4	S.H
11 Mar 2025	8 : 30	3.5	3.2	6.0	S.H
12 Mar 2025	9 : 30	3.5	3.2	6.5	S.H
13 Mar 2025	12 : 00	3.2	3.0	6.0	S.H
14 Mar 2025	16 : 00	3.4	3.0	6.8	R.Y
15 Mar 2025	8 : 30	3.4	3.0	5.9	S.H
16 Mar 2025	:				
17 Mar 2025	9 : 00	3.2	3.0	6.2	S.H
18 Mar 2025	9 : 00	3.2	3.0	6.1	S.H
19 Mar 2025	8 : 40	3.1	3.0	5.9	Y.N
20 Mar 2025	:				
21 Mar 2025	8 : 30	3.3	3.0	6.0	S.H
22 Mar 2025	9 : 00	3.5	3.1	6.0	S.H
23 Mar 2025	:				
24 Mar 2025	14 : 00	3.3	3.1	6.3	S.H
25 Mar 2025	9 : 00	3.4	3.1	6.3	S.H
26 Mar 2025	9 : 00	3.5	3.0	6.3	S.H
27 Mar 2025	9 : 50	3.5	3.1	6.3	Y.N
28 Mar 2025	9 : 50	3.3	3.1	6.5	S.H
29 Mar 2025	8 : 30	3.5	3.0	6.5	S.H
30 Mar 2025	:				
31 Mar 2025	8 : 30	3.5	3.0	6.4	S.H

Investigational Product Manager's Signature: Shunsuke Hitosugi

Date: 01/Apr/2025