

# Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: ER-49C

Thermometer Device/ID: FS Traceable/230735256(Expiry date : 01 Nov 2025)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 May 2025	:				
02 May 2025	14 : 00	3.4	2.9	6.4	S.H
03 May 2025	:				
04 May 2025	:				
05 May 2025	:				
06 May 2025	:				
07 May 2025	8 : 45	3.2	3.1	6.3	S.H
08 May 2025	10 : 40	3.1	3.1	6.6	S.H
09 May 2025	9 : 30	3.6	3.0	6.2	S.H
10 May 2025	9 : 15	3.3	3.1	6.2	S.H
11 May 2025	:				
12 May 2025	8 : 30	3.3	3.1	6.4	S.H
13 May 2025	9 : 30	3.4	3.1	6.2	S.H
14 May 2025	9 : 35	3.5	3.0	6.2	R.Y
15 May 2025	10 : 20	3.3	3.1	6.2	Y.N
16 May 2025	9 : 00	3.5	3.1	6.3	S.H
17 May 2025	8 : 30	3.5	3.2	6.4	S.H
18 May 2025	:				
19 May 2025	9 : 00	3.7	3.1	6.6	S.H
20 May 2025	9 : 30	3.5	3.4	6.2	Y.N
21 May 2025	16 : 00	3.6	3.2	6.4	Y.N
22 May 2025	9 : 30	3.7	3.5	6.4	S.H
23 May 2025	9 : 30	3.5	3.4	6.3	S.H
24 May 2025	8 : 30	3.6	3.2	6.1	S.H
25 May 2025	:				
26 May 2025	9 : 00	3.3	3.1	6.3	S.H
27 May 2025	8 : 30	3.3	3.1	6.1	S.H
28 May 2025	9 : 00	3.5	3.1	6.2	S.H
29 May 2025	12 : 00	3.4	3.1	6.3	Y.N
30 May 2025	9 : 30	3.5	3.1	6.3	S.H
31 May 2025	8 : 30	3.4	3.1	6.2	S.H

Investigational Product Manager's Signature: Shunsuke Hitosugi

Date: 05/June/2025