

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: MPR-514-PJ

Thermometer Device/ID: FS Traceable/210183321 (Expiry date : 09 Mar 2023)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Jan 2023	:				
02 Jan 2023	:				
03 Jan 2023	:				
04 Jan 2023	8 : 30	4.9	4.5	5.0	S-H
05 Jan 2023	8 : 30	4.9	4.6	5.6	Y, N
06 Jan 2023	10 : 00	5.1	4.6	5.5	S-H
07 Jan 2023	9 : 00	4.8	4.3	5.1	S-H
08 Jan 2023	:				
09 Jan 2023	:				
10 Jan 2023	10 : 00	4.9	4.7	5.1	S-H
11 Jan 2023	10 : 30	5.0	4.8	5.1	S-H
12 Jan 2023	9 : 00	5.0	4.9	5.1	S-H
13 Jan 2023	11 : 30	5.0	4.7	5.1	R-Y
14 Jan 2023	9 : 00	4.9	4.7	5.0	S-H
15 Jan 2023	:				
16 Jan 2023	9 : 15	4.8	3.9	5.1	S-H
17 Jan 2023	14 : 00	5.0	4.7	5.1	S-H
18 Jan 2023	15 : 00	4.8	4.7	5.1	S-H
19 Jan 2023	9 : 00	4.9	4.7	5.1	S-H
20 Jan 2023	9 : 00	4.8	4.6	5.0	S-H
21 Jan 2023	9 : 30	4.8	4.6	5.0	Y, N
22 Jan 2023	:				
23 Jan 2023	9 : 30	4.8	4.6	5.0	S-H
24 Jan 2023	9 : 30	4.8	4.6	5.0	S-H
25 Jan 2023	14 : 30	4.7	4.6	4.9	S-H
26 Jan 2023	8 : 35	5.1	4.8	5.2	R-Y
27 Jan 2023	9 : 00	4.9	4.7	5.2	S-H
28 Jan 2023	9 : 00	4.8	4.7	5.2	S-H
29 Jan 2023	9 : 30				
30 Jan 2023	9 : 00	4.8	4.7	5.2	S-H
31 Jan 2023	10 : 00	4.9	4.7	5.1	S-H

Xerror
28-Jan/2023
S-H