

# Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: MPR-514-PJ

Thermometer Device/ID: FS Traceable/221808080(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Feb 2023	9 : 30	4.4	4.4	5.2	S-H
02 Feb 2023	11 : 00	4.7	4.2	4.9	S-H
03 Feb 2023	9 : 00	4.5	4.0	4.9	S-H
04 Feb 2023	9 : 30	4.7	4.1	5.0	R-Y
05 Feb 2023	:				
06 Feb 2023	9 : 00	4.8	4.3	5.0	R-Y
07 Feb 2023	9 : 30	4.9	4.3	5.0	R-Y
08 Feb 2023	9 : 30	4.4	4.3	4.9	Y.N
09 Feb 2023	9 : 00	4.5	4.2	5.0	Y.N
10 Feb 2023	9 : 00	4.7	4.2	4.9	Y.N
11 Feb 2023	:				
12 Feb 2023	:				
13 Feb 2023	9 : 00	4.3	4.2	4.9	Y.N
14 Feb 2023	9 : 00	4.6	4.2	4.9	Y.N
15 Feb 2023	9 : 00	4.7	4.2	4.9	Y.N
16 Feb 2023	8 : 30	4.6	4.1	5.0	Y.N
17 Feb 2023	9 : 00	4.9	4.1	4.8	Y.N
18 Feb 2023	9 : 00	4.4	4.1	4.9	
19 Feb 2023	:				
20 Feb 2023	9 : 00	4.8	4.2	4.8	Y.N
21 Feb 2023	9 : 00	4.3	4.2	5.0	Y.N
22 Feb 2023	11 : 00	4.4	4.2	5.1	Y.N
23 Feb 2023	<del>9 : 00*</del>				
24 Feb 2023	9 : 00	4.5	4.3	5.2	Y.N
25 Feb 2023	9 : 30	4.4	4.2	5.0	Y.N
26 Feb 2023	:				
27 Feb 2023	9 : 30	4.3	4.2	4.9	R-Y
28 Feb 2023	10 : 25	4.6	4.1	4.8	R-Y

\* error  
Yumi Nakamura  
24-Feb-2023