

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: MPR-514-PJ

Thermometer Device/ID: FS Traceable/221808080(Expiry date : 05 Oct 2024)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Jul 2024	9 : 00	5.4	5.3	5.5	S.H
02 Jul 2024	10 : 00	5.4	5.3	5.6	S.H
03 Jul 2024	13 : 00	5.4	5.3	5.5	S.H
04 Jul 2024	8 : 30	5.4	5.3	5.8	S.H
05 Jul 2024	8 : 30	5.2	5.2	5.8	S.H
06 Jul 2024	8 : 30	5.2	5.2	5.4	S.H
07 Jul 2024	:				
08 Jul 2024	8 : 30	5.2	5.2	5.4	S.H
09 Jul 2024	10 : 25	5.5	5.1	5.5	R.Y
10 Jul 2024	9 : 05	5.5	5.1	5.5	Y.N
11 Jul 2024	9 : 25	5.2	5.1	5.5	R.Y
12 Jul 2024	9 : 20	5.4	5.1	5.4	Y.N
13 Jul 2024	8 : 30	5.2	5.1	5.4	Y.N
14 Jul 2024	:				
15 Jul 2024	:				
16 Jul 2024	9 : 00	5.3	5.1	5.4	R.Y
17 Jul 2024	9 : 20	5.2	5.1	5.4	Y.N
18 Jul 2024	9 : 00	5.3	5.1	5.4	S.H
19 Jul 2024	12 : 00	5.2	5.0	5.5	S.H
20 Jul 2024	9 : 00	5.2	5.0	5.5	S.H
21 Jul 2024	:				
22 Jul 2024	8 : 30	5.2	5.0	5.3	S.H
23 Jul 2024	9 : 15	5.2	5.0	5.3	S.H
24 Jul 2024	8 : 30	5.1	5.0	5.5	S.H
25 Jul 2024	8 : 35	5.2	5.0	5.3	R.Y
26 Jul 2024	9 : 00	5.3	5.0	5.4	S.H
27 Jul 2024	8 : 30	5.3	5.0	5.4	S.H
28 Jul 2024	:				
29 Jul 2024	8 : 30	5.1	5.1	5.3	S.H
30 Jul 2024	8 : 30	5.2	5.1	5.3	S.H
31 Jul 2024	11 : 30	5.2	5.0	5.4	S.H

Investigational Product Manager's Signature:

Shunsuke Hitosugi

Date: 03/Aug/2024