

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: MPR-514-PJ

Thermometer Device/ID: FS Traceable/230735257(Expiry date : 01 Nov 2025)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Feb 2025	9 : 30	5.6	3.7	5.7	S.H
02 Feb 2025	:				
03 Feb 2025	9 : 00	5.1	3.7	5.7	S.H
04 Feb 2025	9 : 30	5.4	3.8	5.8	S.H
05 Feb 2025	9 : 00	5.2	3.8	5.8	S.H
06 Feb 2025	14 : 00	5.4	3.5	5.7	Y.N
07 Feb 2025	9 : 00	4.5	3.5	5.7	S.H
08 Feb 2025	10 : 00	5.6	3.7	6.5	R.Y
09 Feb 2025	:				
10 Feb 2025	9 : 30	4.2	3.6	6.0	S.H
11 Feb 2025	:				
12 Feb 2025	8 : 50	5.7	3.6	6.0	Y.N
13 Feb 2025	9 : 00	4.9	3.6	6.0	S.H
14 Feb 2025	8 : 30	5.8	3.6	5.9	Y.N
15 Feb 2025	9 : 00	5.7	3.8	5.8	S.H
16 Feb 2025	:				
17 Feb 2025	9 : 00	5.8	3.8	5.8	S.H
18 Feb 2025	9 : 30	4.2	3.5	6.0	S.H
19 Feb 2025	9 : 00	4.0	3.9	5.9	S.H
20 Feb 2025	9 : 00	5.7	3.9	5.9	S.H
21 Feb 2025	12 : 00	4.2	3.7	6.1	Y.N
22 Feb 2025	11 : 30	4.6	3.7	6.1	Y.N
23 Feb 2025	:				
24 Feb 2025	:				
25 Feb 2025	8 : 30	5.1	4.0	6.0	Y.N
26 Feb 2025	9 : 30	5.9	4.0	6.0	S.H
27 Feb 2025	10 : 00	4.3	3.9	6.1	S.H
28 Feb 2025	9 : 30	5.3	4.2	6.0	S.H

Investigational Product Manager's Signature: Shunsuke Hitosugi

Date: 03/Mar/2025