

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: MPR-514-PJ

Thermometer Device/ID: FS Traceable/230735257(Expiry date : 01 Nov 2025)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Jun 2025	:				
02 Jun 2025	8 : 40	4.5	4.0	5.8	Y.N
03 Jun 2025	11 : 30	5.3	4.0	5.8	S.H
04 Jun 2025	9 : 10	4.0	3.9	6.0	R.Y
05 Jun 2025	9 : 00	5.6	3.8	5.8	S.H
06 Jun 2025	9 : 00	3.9	3.8	6.0	S.H
07 Jun 2025	9 : 40	3.2	3.1	6.3	Y.N
08 Jun 2025	:				
09 Jun 2025	9 : 00	4.1	3.8	6.0	R.Y
10 Jun 2025	9 : 30	4.6	3.6	6.0	S.H
11 Jun 2025	9 : 20	5.6	3.8	6.1	S.H
12 Jun 2025	10 : 00	5.4	3.9	6.5	S.H
13 Jun 2025	9 : 20	4.9	3.9	6.5	S.H
14 Jun 2025	9 : 00	5.5	4.4	5.9	S.H
15 Jun 2025	:				
16 Jun 2025	9 : 00	5.0	4.9	5.9	S.H
17 Jun 2025	9 : 25	4.5	4.1	6.2	R.Y
18 Jun 2025	9 : 10	5.7	4.1	5.9	S.H
19 Jun 2025	9 : 00	5.5	4.1	6.0	S.H
20 Jun 2025	9 : 00	4.4	4.2	5.9	S.H
21 Jun 2025	9 : 30	5.2	4.2	5.9	S.H
22 Jun 2025	:				
23 Jun 2025	8 : 30	5.8	4.2	6.0	S.H
24 Jun 2025	10 : 00	5.8	4.2	6.0	S.H
25 Jun 2025	11 : 00	5.5	4.2	6.0	R.Y
26 Jun 2025	14 : 30	5.3	4.2	6.0	Y.N
27 Jun 2025	9 : 00	4.4	4.2	6.0	S.H
28 Jun 2025	9 : 30	4.6	4.2	6.0	S.H
29 Jun 2025	:				
30 Jun 2025	8 : 30	4.7	4.2	6.0	S.H

Investigational Product Manager's Signature:

Shunsuke Hitosugi

Date: 03-Jul-2025