

# Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : 2°C to 8°C

Refrigerator/Freezer Name: MPR-514-PJ

Thermometer Device/ID: FS Traceable/230735257(Expiry date : 01 Nov 2025)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 Sep 2025	9 : 15	5.6	3.8	6.1	S.H
02 Sep 2025	10 : 40	4.8	4.0	6.0	S.H
03 Sep 2025	15 : 00	5.3	3.8	6.0	S.H
04 Sep 2025	9 : 30	5.9	4.2	6.0	Y.N
05 Sep 2025	9 : 30	5.3	4.2	6.0	S.H
06 Sep 2025	8 : 30	5.2	4.7	6.0	S.H
07 Sep 2025	:				
08 Sep 2025	8 : 30	4.2	4.1	6.0	Y.N
09 Sep 2025	9 : 40	4.4	4.1	6.0	Y.N
10 Sep 2025	10 : 30	5.9	4.1	6.0	P.Y
11 Sep 2025	9 : 30	4.3	4.1	6.2	S.H
12 Sep 2025	13 : 00	4.2	3.6	6.0	S.H
13 Sep 2025	9 : 00	4.6	4.0	6.0	S.H
14 Sep 2025	:				
15 Sep 2025	:				
16 Sep 2025	9 : 50	5.9	4.0	6.0	S.H
17 Sep 2025	9 : 20	4.1	4.0	6.0	S.H
18 Sep 2025	9 : 30	4.8	4.0	5.2	S.H
19 Sep 2025	10 : 40	4.4	4.0	6.3	S.H
20 Sep 2025	10 : 00	5.4	4.5	6.0	S.H
21 Sep 2025	:				
22 Sep 2025	8 : 30	4.1	4.0	6.0	S.H
23 Sep 2025	:				
24 Sep 2025	8 : 30	5.0	4.0	5.9	S.H
25 Sep 2025	15 : 30	4.2	3.7	6.2	S.H
26 Sep 2025	9 : 00	4.1	3.6	6.2	S.H
27 Sep 2025	9 : 00	4.5	3.8	6.2	S.H
28 Sep 2025	:				
29 Sep 2025	9 : 15	4.7	3.8	6.2	S.H
30 Sep 2025	9 : 00	4.2	3.9	6.2	Y.N

Investigational Product Manager's Signature:

Shunsuke Hitosugi

Date: 03/Oct/2025