

Temperature Log

The Jikei University Hospital

Location : Pharmaceutical department

Acceptable Temperature Range : room temperature

Thermometer Device/ID: FS Traceable /230735258(Expiry date 01 Nov 2025)

Date 日付	Time 確認時間	Current 確認時の 温度	Min 最低温度	Max 最高温度	Initials (確認者)
01 May 2025	:				
02 May 2025	14 : 00	21.0	20.6	21.6	S.H
03 May 2025	:				
04 May 2025	:				
05 May 2025	:				
06 May 2025	:				
07 May 2025	8 : 45	20.9	20.4	21.6	S.H
08 May 2025	10 : 40	21.0	20.6	21.6	S.H
09 May 2025	9 : 30	20.9	20.7	21.5	S.H
10 May 2025	9 : 15	21.0	20.6	21.5	S.H
11 May 2025	:				
12 May 2025	8 : 30	21.0	20.6	21.5	S.H
13 May 2025	9 : 30	20.9	20.5	21.5	S.H
14 May 2025	9 : 35	21.2	20.6	21.6	R.Y
15 May 2025	10 : 20	21.0	20.7	21.5	Y.N
16 May 2025	9 : 00	21.1	20.0	21.6	S.H
17 May 2025	8 : 30	21.4	20.6	21.7	S.H
18 May 2025	:				
19 May 2025	9 : 00	21.0	20.5	22.9	S.H
20 May 2025	9 : 30	20.9	20.5	21.4	Y.N
21 May 2025	16 : 00	21.3	20.5	21.5	Y.N
22 May 2025	9 : 30	21.3	20.6	21.7	S.H
23 May 2025	9 : 30	21.1	20.7	21.5	S.H
24 May 2025	8 : 30	21.2	21.0	21.4	S.H
25 May 2025	:				
26 May 2025	9 : 00	21.1	20.9	21.5	S.H
27 May 2025	9 : 00	21.2	21.0	21.5	S.H
28 May 2025	8 : 30	21.0	20.8	21.5	S.H
29 May 2025	12 : 00	21.1	20.9	21.5	Y.N
30 May 2025	9 : 30	21.3	20.9	21.6	S.H
31 May 2025	8 : 30	21.1	20.9	21.5	S.H

Investigational Product Manager's Signature: Shimizu Hitosugi

Date: 05/Jun/2025